

Centrifuge Tube

Date 11112022

Form # 1190

Part # 002461

This is to certify that the Magnaflux Centrifuge Tube /100ML meets the following requirements:

- ASME Boiler & Pressure Vessel Code, Sec. V, 2021 Edition, Article 7 T-765
- ASTM E709-21, Paragraph 20.6.1, Appendixes X5
- ASTM E 1444-22, Paragraph 7.2.1.1

All Volumetric standards used for calibration are traceable to the National Institute of Standards and Technology with the following accuracy.

0 to 1.5 ML	+/- .03 ML
above 1.5 to 3 ML	+/- .20 ML
above 3 to 5 ML	+/- .30 ML
above 5 to 10 ML	+/- .5 ML
above 10 to 25 ML	+/- 1 ML
above 25 ML	+/- 2 ML



Roger St. Claire
Quality Control Manager