

# Centrifuge Tube

**Date 11112022**

**Form # 1180**

**Part # 8493**

This is to certify that the Magnaflux Centrifuge Tube /100ML meets the following requirements:

- ASME Boiler & Pressure Vessel Code, Sec. V, 2021 Edition, Article 7 T-765
- ASTM E709-21, Paragraph 20.6.1, Appendixes X5
- ASTM E 1444-22, Paragraph 7.2.1.1

All Volumetric standards used for calibration are traceable to the National Institute of Standards and Technology with the following accuracy.

0-1.0 mL	$\pm 0.02$ mL
2 mL	$\pm 0.20$ mL
Above 2 mL to 10 mL	$\pm 0.50$ mL
Above 10 mL to 25 mL	$\pm 1$ mL
Above 25 mL to 100 mL	$\pm 2$ mL



Roger St. Claire  
Quality Control Manager